

THE 67<sup>TH</sup> INTERNATIONAL



# OPEN READINGS

CONFERENCE FOR STUDENTS OF PHYSICS AND NATURAL SCIENCES

**BOOK OF  
ABSTRACTS** | **2024**



Vilnius  
University

VILNIUS UNIVERSITY PRESS

Editors:

Martynas Keršys  
Rimantas Naina  
Vincentas Adomaitis  
Emilijus Maskvytis

Cover and Interior Design:

Goda Grybauskaitė

Vilnius University Press  
9 Saulėtekio Av., III Building, LT-10222 Vilnius  
info@leidykla.vu.lt, [www.leidykla.vu.lt/en/](http://www.leidykla.vu.lt/en/)  
[www.knygynas.vu.lt](http://www.knygynas.vu.lt), [www.journals.vu.lt](http://www.journals.vu.lt)

Bibliographic information is available  
on the Lithuanian Integral Library Information System (LIBIS) portal [www.ibiblioteka.lt](http://www.ibiblioteka.lt)  
ISBN 978-609-07-1051-7 (PDF)

© Vilnius University, 2024

# ELECTROCHEMICAL IMMUNOSENSOR BASED ON GOLD NANOSTRUCTURES FOR THE DETECTION OF ANTIBODIES AGAINST SARS-COV-2 SPIKE PROTEIN

Kristina Sobol<sup>1</sup>, Katazyna Blazevic<sup>1</sup>, Benediktas Brasiunas<sup>1</sup>, Almira Ramanaviciene<sup>1</sup>

<sup>1</sup> NanoTechnas - Center of Nanotechnology and Materials Science, Institute of Chemistry, Faculty of Chemistry and Geosciences, Vilnius University  
[kristina.sobol@chgf.stud.vu.lt](mailto:kristina.sobol@chgf.stud.vu.lt)

Electrochemical immunosensors have become fundamental tools in various fields, including biomedical diagnostics, disease progression monitoring, environmental surveillance, food quality control, forensic analysis, and biomedical research. These highly specific analytical systems demonstrate exceptional sensitivity and efficiency, enabling precise identification of analyte presence and its concentration in a sample [1].

The principle operation of an electrochemical immunosensor is based on the electrochemical signal recording after the specific interaction between immobilized antigen and antibodies present in the solution. Essential requirements for immunosensors are their selectivity, sensitivity, repeatability, accuracy, and reproducibility. To enhance the sensitivity of the immunosensor, the surface of the electrode is frequently modified with nanomaterials, known for their unique physical, chemical, and optical properties. Compared to traditional immunosensors, nanomaterial-based immunosensors are more specific and sensitive, allowing the detection of lower analyte concentration [2].

In this study, an electrochemical immunosensor was developed for investigation of the interaction between immobilized SARS-CoV-2 virus S protein and specific monoclonal antibodies. The impact of various gold nanostructures formed on the electrode on the performance of the immunosensor was evaluated. Additionally, the influence of non-specific interactions on the analytical signal was investigated, along with methods used to reduce such effect. This research contributes valuable insights into the optimization and application of electrochemical immunosensors for biomedical diagnostics.

## Acknowledgement

This research was funded by a grant (No. S-MIP-22-46) from the Research Council of Lithuania.

- 
- [1] M. Zembala, Z. Adamczyk, Measurements of streaming potential for mica covered by colloid particles, *Langmuir* 2000, 16, 1593-1602 (1999).  
 [2] T. Luxbacher, *The ZETA Guide: Principles of the streaming potential technique*. 27-30. (Anton Paar GmbH, 2014).